

# LOW PRESSURE VERIFICATION OF GAS ULTRASONIC FLOW METERS AS AN EFFICIENT SOLUTION TO THE METROLOGICAL CONFIRMATION PROCESS SUGGESTED IN ISO 10012

Henry Abril; Jose A. Fuentes; Jorge Reyes

Corporación CDT de GAS

km 2 Vía Refugio, Piedecuesta - Colombia

[habril@cdtdegas.com](mailto:habril@cdtdegas.com); [jfuentes@cdtdegas.com](mailto:jfuentes@cdtdegas.com);  
[jreyes@cdtdegas.com](mailto:jreyes@cdtdegas.com)

Juan M. Ortiz; John F. Velosa

Transportadora de Gas Internacional S.A. ESP

Cra 34 No. 41-51 Bucaramanga - Colombia

[juanmanuel.ortiz@tgi.com.co](mailto:juanmanuel.ortiz@tgi.com.co) ; [john.velosa@tgi.com.co](mailto:john.velosa@tgi.com.co)

## Abstract

The metrological confirmation process applied to gas ultrasonic flow meters, according to the terms specified by the technical standard ISO 10012, represents a high challenge for natural gas transmission companies which, as TGI, are interested in establishing a Measurement Management System based on such standard.

In this paper, a metrological control proposal is described for exclusive application to ultrasonic flow meters (USMs) operating at low (atmospheric) pressures. This is based on tests performed on 2 new flow meters DN150 (6") which were calibrated at first in PIGSAR (Germany) using natural gas at high pressure, as test fluid, and later characterized by air at atmospheric pressure in CDT de GAS (Colombia).

Details of performed tests and obtained results are presented as well as the corresponding technical comments and observations. Based on the analysis of the results, a verification proposal is conformed according to the guidelines for metrological confirmation process suggested by ISO 10012. The decision-taking processes are highlighted and supported by control limits in order to assure the veracity, validity and reliability of measurements. Finally, a section about the application prospects, considering technical and economic issues, is offered before the general conclusions and recommendations of this work.

## 1. Introduction

Measurements carried out in each custody transfer point (reception or delivery), are one of the key processes for fulfilling the social purpose of natural gas transport companies, because such measurements allow controlling gas balances and invoicing transport services.

In Colombia, the regulations [**Error! No se encuentra el origen de la referencia.**] dealing with measurement processes do not conform to OIML R140 [2] in matters of the performance specifications require by the elements making up a measurement system, however, such regulation do state an overall performance applied to measurement systems as well as a control limit for gas balances of transport network. Taking into account these aspects the metrological assurance for correlated magnitude, such as pressure, temperature, volume and energy, is defined.

Carrying out such metrological assurance, according to standards established at worldwide, implies making an investment at a different level of the process which cover both installed elements and results obtained by measurement processes.

In Colombia, since 2007, Transportadora de Gas Internacional (TGI) S.A., along with the technological support of CDT de GAS Corporation, has gradually used a measurement management system based on international standard ISO 10012, which allows improving gas balances and keeping it under control (328 measurement system of transfer custody in a 3957-kilometer pipeline system), in order to obtain the best cost/benefit ratio from a world perspective.

## 2. Metrological Confirmation

The use of a metrological confirmation program is technically and economically feasible in a local background for magnitudes such as pressure, temperature and energy, considering both metrological infrastructure and mechanisms for its implementation.

Nevertheless, in the particular case of measurement of gas volume, using a metrological confirmation program is constrained due to factors such as:

- a. Scarce availability of calibration laboratories for ranges of flow rate and pressure such as those used in transport pipeline systems. In Colombia, calibration laboratories

- operate using air at atmospheric pressure and just one laboratory uses natural gas, at approximately 17 Bar.
- b. Logistics aspects required for moving a meter from its location to a laboratory.
  - c. Need for appealing to reconciliation methods due to absence of a back meter during a calibration process.
  - d. Investments required and maintenance costs of the model once implemented.

### **2.1. Usual practice for metrological confirmation of volume meters**

In the case of some measurement technologies used in custody transfer, such as rotary, turbine and Coriolis meters, the troubles described above may be overcome relatively easily, considering the laboratories structure allowing each element's performance evaluation in a proper way, according to standard applicable to each measurement technology.

Nonetheless, in the case of ultrasonic meters (USM), there is no agreement related to a methodology applied to its metrological confirmation. This can be evidenced for the absence of guidelines or recommendations that allow setting such calibration intervals. Experimental works presented by C. Coull and E. Spearman **¡Error! No se encuentra el origen de la referencia.** for example, attempted to determine USM performance with time through periodical calibrations. Those studies, however, were not conclusive as for the sources generating the deviations found and the possibility to predict other meters' performance.

On the other hand, some works have achieved to generate metrological confirmation strategies for USMs widely used in some sectors of industry, such as Peterson's strategy - Lightbody [4], which is mainly applied in the United Kingdom. Such strategy considers that a new USM must be calibrated six months after being placed in operation, then, six more months later, a second calibration must be carried out. In that time, if the difference FWME were less than 0,3% (between two subsequent calibrations), such USM would be qualified for a yearly calibration. Additionally to Peterson's strategy, the possibility of using USM diagnostic tools as a means of extending its calibration intervals is posed.

T. Kegel and S. English [3] developed an interesting predictive mathematical model to estimate the possible meter's deviation with time, which would ease the estimation of USMs recalibration periods using as inputs: meter size, working fluid mean velocity and the current recalibration period. This was done since they found no specification of such intervals in any AGA 9 edition.

Finally, it is important to mention that developments on ultrasonic measurement, such as the usage of transducers for coupling low pressure wave, have opened the possibility of carrying out USM metrological confirmations, using air at atmospheric pressure as stated by [5]. This could lead to estimate that meter calibrations at low pressure may

substitute high pressure calibrations, at least partially, particularly for recalibration processes, taking into account, of course, comparability between results at high and low pressure; and the economic advantages implied.

### **2.2. TGI's Strategy for ISO 10012 implementation, associated to measurement of gas volume:**

Taking into account the wide range of measurement systems in the TGI's transport network, defining a unified strategy, applicable to its entire infrastructure, is not possible. This can be observed by means of a Pareto analysis from measured volume by each meter in relation to the total volume transported by TGI.

- A-type Systems: 26 measurement systems quantifying 80% out of the total transported gas. Such meters are mainly ultrasonic meters and at a smaller extent, turbine flow meters.
- B-type and C-type Systems: 302 measurement systems quantify 20% out of the total transported gas. These meters are mainly rotary meters and turbine flow meters and some Coriolis meters.

From the metrological confirmation practice previously described and the distribution of measurement systems according to their capacity, two scenarios for using of metrological confirmation were defined, each of them has its own particular strategy.

#### **A-type Systems**

This strategy consists of taking advantage of technological updating process of measurement systems in order to implement USM whose technology allows verifications at atmospheric pressure. This allows taking good profit of the calibration facilities available locally.

#### **B-type and C-type Systems**

This strategy consisted in developing a mobile calibration facility able to carry out metrological calibration processes in-situ. Such facility named M3Tlab [6] was developed by TGI in conjunction with CDT de GAS and is currently in a pre-operative stage in order to validate its performance and obtain accreditation ISO 17025.

### **3. Metrological confirmation strategy for high impact systems**

In this section, a strategy for carrying out a metrological confirmation processes exclusive for USMs is described. Such strategy is mainly based on ultrasonic meter ability of coupling signals from transducers at low pressure, thus they can be assessed metrologically, using facilities that operate at atmospheric pressure, minimizing the effect of infrastructure limitations and metrological assurance costs. In the Figure 1. the different stages, that the strategy has, are showed. The

metrological confirmation process of USMs starts as soon as it is acquired.

### Stage 1

An initial calibration is carried out. This initial process includes both a calibration at estimated operation conditions (Between 17 and 70 bars, using natural gas) and a calibration at atmospheric pressure to evaluate the comparability of results at different conditions.

### Stage 2

This stage is composed by diagnostic activities (by means of auto-diagnostic capacities) and inspection are periodically executed (every month) with the objective of evidencing a correct performance of the meter, acceptable process conditions and performance parameters according to the allowable boundaries (for instance SNR, Performance, SOS). If the obtained results are out of the limits, verification activities at low pressure are necessary.


### Stage 3

Verification activities represent the most important part of the strategy; it consists of calibration of USM at atmospheric pressure condition, using air as calibration fluid, and compares such results to initial ones. If results do not have an acceptable equivalent degree and especial considerations, which support the hypothesis, have been controlled during the verification process, recalibration activities at high pressure is need. The verification activity is carried out when is reached a period of five years or when a deviation is detected in performance parameters of USM.

### Stage 4

Finally, recalibration activity at high pressure using natural gas as calibration fluid will just be made when a considerable deviation out of the limits is detected in the verification stage.

Figure 1. Metrologic Confirmation Strategy applied to a USM



The usage of the strategy described above is based on the confirmation of two hypotheses related to results equivalency.

- **Hypothesis 1:** If the results obtained by USM calibrations at high and low pressure, using natural gas and air as calibration fluid respectively, have an acceptable equivalent degree, a later proof at low pressure can be used as a tool to confirm if USM performance agrees to results of high pressure calibration.
- **Hypothesis 2:** If the results of verification process, by means of periodic calibrations at low pressures, have an acceptable equivalent degree, such results may be used as support for extending the validity of calibration time at high pressure, considering possible factors of influence in the applied method.

Taking into account the involved variables and different factor that may influence results obtained by the verification process, the following considerations and clarifications for the development of the described hypotheses were defined, in order to

- The laboratory, which carried out the calibration at low pressure using air as calibration fluid, is recognized under the guidelines given by ISO 17025 and has developed proficiency tests.
- The stability over time and reproducibility of the calibration bench were included as part of the sources of uncertainty, therefore limits used for assessing the equivalency of the results with the calibration bench, consider the effect of a likely data dispersion for such aspects.
- In the evaluation of the results described below was used the mathematic model of normalized error as a statistical performance indicator of results on time.
- The methodology used in the verification process, both initial and posteriors, contemplates the control of extern factors which can affect the meter performance

## 4. Results


In order to validate the implementation of the strategy, tests were carried out and evidences from the meter's performance were obtained using a DN150 (6") meter. This device was acquired in 2011 in Germany and was calibrated at high pressure in PIGSAR and, as soon as it arrived in CDT de GAS facilities, before commissioning, it was calibrated at atmospheric pressure using air as the working fluid. In February 2013 the meter was tested again, running at atmospheric pressure.

### 4.1. Equivalence of results at high and low pressure: initial calibrations in laboratories

The initial calibration was carried out at operating conditions (high pressure 13, 79 bar -200 psig) in PIGSAR Germany.

Both the calibration and atmospheric pressure characterization were performed using upstream and downstream pipelines in order to mitigate the effects of “different installation conditions” and the possible misalignment effects described in [4]. During the calibration and characterization processes the self-diagnosis reports necessary for its use as base line were issued as well as the results of calibration and characterization at low pressure, which allow showing further deviations. The results of the calibrations of the meter under evaluation at natural gas pressure and the calibration with air at atmospheric pressure (acquired in 2011), are shown below (figure 2).

**Figure 2.** PIGSAR calibration curves (high pressure) and CDT de GAS (atmospheric pressure)



In order to evaluate the equivalence of the results obtained with a meter at different operating conditions, a comparative analysis between calibration results obtained by PIGSAR and CDT de GAS was performed

The standard error in the comparative analysis is used. This error is estimated for each flow rate and represents the equivalence degree between the two conditions.

$$E_n = \frac{X_{CDT\ de\ GAS} - X_{PIGSAR}}{\sqrt{U_{CDT\ de\ GAS}^2 + U_{PIGSAR}^2}} \quad \text{Equation 1}$$

Where,

- $X_{CDT\ de\ GAS}$  Percentage error – CDT de GAS laboratory – low pressure
- $X_{PIGSAR}$  Percentage error – PIGSAR – high pressure
- $U_{CDT\ de\ GAS}$  Expanded uncertainty of calibration reported by CDT de GAS
- $U_{PIGSAR}$  Expanded uncertainty of calibration reported by PIGSAR

The result is evaluated by using the following criterion:

$E_n < 1$	The result is considered comparable
$1 < E_n < 1,2$	The result is considered comparable with cautions
$E_n > 1,2$	The result is considered non-comparable

It is expected to use this criterion in the evaluation to demonstrate the meter’s ability to provide a similar performance at high and low pressures. If the results happen to be **non-comparable**, it would mean that there are significant differences for working at high and low pressures or that there may be deviations between the calibration facilities, therefore this condition does not provide conclusive supports to take decisions and the current strategy would not be applicable.


**Table 1.** Standard errors calculated for comparison between high and low pressure

Ref.	Caudal $m^3/h$	CDT DE GAS - 2011		
		$X_{CDT} - X_{PIGSAR}$ %	$U_{[CDT,PIGSAR]}$ %	$E_N$ %
PIGSAR	32	0,46	0,87	0,53
	80	-0,43	0,51	-0,85
	238	-0,07	0,27	-0,26
	393	-0,01	0,23	-0,03
	641	0,11	0,24	0,46
	1126	-0,03	0,23	-0,14
	1611	0,11	0,34	0,33

Table 1 shows the standard errors for the meter under evaluation, being all of them less than one (1), which means that, for this specific case, there is an equivalence between the reported calibrations (validating the first hypothesis) at low pressure using atmospheric air and high pressure using natural gas..

This result is shown graphically (figure 3) by using the value of the dividend of equation 1 (total uncertainty) as the permissible limit and the value to be evaluated is the difference between percentage errors reported by each laboratory.

**Figure 3.** Curve representing the error difference between CDT and PIGSAR




#### 4.2. Results of verifications in CDT de GAS laboratory.

The USM verification is performed by carrying out tests in the low-pressure atmospheric-air calibration facilities at

CDT de GAS. The meter’s evaluation using this criterion pretends to evidence the robustness of the verification process with time. In case that a **non-comparable** result is issued, the methodology cannot be applied.

**Figure 4.** Initial and latter calibration curve at CDT de GAS – atmospheric pressure



The results shown correspond the meter 1, between the initial verification performed in 2011 and a subsequent verification in February 2013 (See Figure 4).

In order to evaluate the equivalence between the initial and latter calibration in a period of time, the evaluation was performed similarly as it is carried out to evaluate at different pressures, using the algorithm of standard error:

$$En = \frac{X_{inicial} - X_{posterior}}{\sqrt{U_{inicial}^2 + U_{posterior}^2}} \quad \text{Equation 2}$$


Where,

- $I_{inicial}$  Percentage error – CDT de GAS laboratory – initial verification
- $X_{posterior}$  Percentage error – CDT de GAS laboratory – latter calibration
- $U_{inicial}$  Expanded uncertainty of initial calibration reported by CDT de GAS
- $U_{posterior}$  Expanded uncertainty of latter calibration reported by CDT de GAS

**Table 2.** Standard errors calculated between initial and latter verification using air at atmospheric pressure

Ref.	Caudal $m^3/h$	CDT DE GAS - 2013		
		$X_{2013} - X_{2011}$ %	$U_{[2013,2011]}$ %	$E_N$ %
CDT DE GAS 2011	32	0,41	1,08	0,38
	80	0,79	0,84	0,94
	238	0,34	0,37	0,93
	393	0,07	0,30	0,25
	641	-0,06	0,29	-0,20
	1126	0,01	0,29	0,03
	1611	-0,15	0,31	-0,47

**Figure 5.** Error differences curve between CDT initial (2011) and CDT latter (2013).



Similarly, Table 2 and Figure 5 evidence standard errors less than one (1), which means that:

- ✓ The initial verification with air at atmospheric pressure (2011) and the latter verification (2013) are equivalent, therefore it is confirmed that high pressure calibration is still valid.

Based on the results of: diagnoses and initial and latter verification at low pressure, as for this specific case, a first evidence was obtained, confirming a priori that: if deviations out of permissible limits are not present in the results, the meter’s performance will not show a noticeable deviation and thus, the verification period may be extended.

## 5. Conclusions

The metrological confirmation strategy proposed in this document is an alternative providing an acceptable confidence level for the measures, and conforms to the metrological infrastructure available in the region.

As for the specific case of the evaluated meter, it was possible to demonstrate its ability to provide comparable results operating at high pressure using natural gas and atmospheric pressure with air, and this ability can be used to perform low pressure intermediate verifications. This fact allows decreasing the meters’ assurance costs without being detrimental to the quality of the results.

## 6. References

- [1]. Reglamento Único de Transporte RUT - CREG
- [2]. OIML R140 :2007 - Measuring Systems For Gaseous Fuel
- [3]. C. Coull and E. Spearman, “Practical Experiences Of Operating Small Bore (4”) Gas Ultrasonic Meter for Fiscal Measurement on a FPSO”, International North Sea Flow Measurement Workshop, 2008.
- [4]. S. Peterson and C. Lightbody, “On line Condition Based Monitoring of Gas USMs”, International North Sea Flow Measurement Workshop, 2008.
- [5]. V. Herrmann, M. Wehmeier, T. Dietz, R. Kramer and B. Mickan, “A new low pressure calibration facility using 8-path ultrasonic meters as working standards”, 6th ISFFM, 2006.
- [6]. M3Tlab. www.m3tlab.com

