

THE SPECIFICATION, SELECTION AND USE OF LIQUID & GAS FLOW RATE MEASURING DEVICES

Richard Fertell
Proteus Industries Inc.
340 Pioneer Way, Mountain View, California, 94041, USA

Abstract

Description of an international standard that is complied with as part of the CE Mark or certified by factory audit for product safety marks, to ensure all liquid, gas & LPG/LNG flow rate measurement technology specifications are consistent for evaluation and comparison as well as to document and advance the state-of-the-art of flow rate measurement technologies.

The standard will cover all liquid, gas & LPG/LNG flow rate technologies, error sources, fluid and temperature applications, and expected measurement uncertainties.

The standard includes all improvement suggestions from the CFM 2011 presentation participants.

Introduction

Which flow rate measuring technologies are appropriate for a user application? Are all flow rate device manufacturers reporting measurement uncertainty using internationally agreed upon standards?

Problem

In order to select and specify the correct type of flow rate measuring device(s) for an application and to compete in a level playing field, various users, industries and manufacturers would like to know the answers to the previous questions.

Solution

Create an international standard that lists all flow rate measuring technologies, associated error sources, fluid and temperature applications, and expected measurement uncertainties. Update the standard annually to keep up-to-date with advances in technology.

Background

Several industries and government agencies publish liquid flow rate measuring standards specifying a single flow technology to be used in only one application. These standards are not practical for other applications and cause confusion and are misleading for users because manufacturers do not state specifications in terms of measurement uncertainty or the same type of specifications. Different technologies and, sometimes, even the same technologies cannot be compared by users because some manufacturers specify accuracy with a limited definition in footnotes that may only include instantaneous stability (repeatability).

Users need to know how flow measurement works. Flow is a dynamic process with inter-relationships between pulsating fluid, turbulence intensity, viscosity, density and primary and secondary devices. Pressures fluctuate, velocity profiles change, vortices form differently and flow-measuring devices respond differently to these changes.

The International Standard

The international standard will define flow states and related measurement discipline (liquid, solid, gas, LPG/LNG for flow and thermal plasma for Electrical) and terms not found in the VIM, such as laminar and turbulent flow. A section will describe issues that influence measurements: turbulence/cavitation and their sources, temperature effects, non-homogeneous/multi-state/compressible fluids. A summary table will allow quick review of technology performance. A section will describe how each technology works, influences and a measurement uncertainty calculation. A section will describe value-added electronics outputs, specifications and measurement uncertainty contribution. A section will describe calibration method versus real use and include discussions a variety of topics that influence the quality of the unit calibration.

Liquid Flow Terminology and Influences

This section will describe issues that can influence measurements, such as: a) turbulence and its sources, b) cavitation and its sources, c) temperature effects on viscosity, density, device clearances, and electronics performance, d) non-homogenous fluids (multiple viscosities or densities), e) multi-state fluids of gas, liquids, solids, f) compressible fluids, g) closed/open systems and closed/open channels, h) flow ranges from drip/hour to Dam feed-throughs for closed channels and trickle to oceanic for open channels, as well as i) calibration method versus real use. The turbulence topic will include a discussion of the Threshold Pulsation Index (2), the Flow-Rate Correction Factor (3), and Reynold's Numbers (Re) for Laminar, Turbulent and Transition Flow Regions as well as flow profiles based upon plumbing configurations and flow velocity.

The Summary Table

A table shall summarize technologies by use parameters, Table 1. The table terms will be described: a) method of measurement, b) standard temperature range, c) extended temperature ranges, d) best measurement uncertainty, e) fluid compatibility, f) sensitivity to density or viscosity, g) use for liquids, gases, solids, slurries, h) Reynold's Number operating range with corresponding viscosity and density ranges.

Technology	Method	True or calculated	Std Temp	Extended Temp	Flow Range	Best MU	Fluid Compatibility	Liquid, gas, slurries	Sensitivities	Moving Part
Standing Pipe/Weigh Standard	Volume or Mass	True	Ambient	-20C to 90C	0,12 to 38,754 liter/min 1mg/hr to 4,000kg/min	0.025%	Stainless Steel	Liquid, Slurries	Temperature	Scale & Valve
Coriolis	Mass	True	0-100C	-40C to 200C	0,02 g/hr to 75 kg/min to 4100 metric ton/hr	0,1% mass 0,2% volume	Stainless Steel	Liquid, Gas, slurries	Density, temperature	Tube
Turbine	Volume	Calculated	0-100C	-40C to 200C	0,57 to 387,54 liter/min	0,27%	Stainless Steel	Gas, Liquid	Turbulence, flow profile, viscosity, temperature	Rotor
Orifice Plate	Volume	Calculated	0-100C	-20C to 100C	0,57 to 378,54 liter/min	1%	Stainless Steel	Liquid	Turbulence, flow profile, viscosity, temperature	Pressure Sensor
Target	Volume	Calculated	0-70C	0C to 70C	0,57 to 378,54 liter/min	3%	Stainless Steel, Plastic	Liquid	Turbulence, flow profile, viscosity, temperature	Spring
Variable Area (rotameter)	Volume	Calculated	0-70C	0C to 70C	0,57 to 378,54 liter/min	3%	Stainless Steel, Plastic	Gas, Liquid	Turbulence, flow profile, viscosity, temperature	Float
Thermal differential	Volume	Calculated	0-70C	0C to 70C	0,57 to 378,54 liter/min	2%	Stainless Steel	Gas, Liquid	Turbulence, flow profile, viscosity, temperature	
Pressure differential	Volume	Calculated	0-100C	0C to 100C	0,57 to 378,54 liter/min	1%	Stainless Steel	Gas, Liquid	Turbulence, flow profile, viscosity, temperature	Pressure Sensor
Ultrasonic	Volume	Calculated	0-70C	0C to 70C	0,57 to 378,54 liter/min	2%	Non-intrusive	Gas, Liquid	Turbulence, flow profile, viscosity, temperature	
Magnetic	Volume	Calculated	0-70C	0C to 70C	38,75 to 378,54 liter/min	1%	Magnetic Fluids Only	Liquid	Non-magnetic fluid	
Gear	Volume	True	0-75C	0C to 75C	<17,4 litre/min	1%	Stainless Steel	Liquid	Viscosity <2cP	Gears
Vortex shedding	Volume	Calculated	0-75C	0C to 75C	0,57 to 378,54 liter/min	1%	Stainless Steel, Plastic	Gas, Liquid	Turbulence, flow profile, viscosity, temperature	Pressure Sensor

Technologies

This section describes how each technology works, influences and a measurement uncertainty calculation. Coriolis, turbine, orifice plate, target, variable area (rotameter), thermal differential, pressure differential, ultrasonic, magnetic, gear, vortex shedding and other technologies are covered. The Turbine Technology Section is an example.

Turbine Technology

A turbine measures volumetric flow rate based upon the rotational speed of a bladed wheel. The rotating wheel is sensitive to the viscosity of the fluid because the force of the fluid transferred to the wheel is dependent upon how much of the fluid slips by the blades. A low viscosity liquid will not transfer as much fluid force to the blades of the wheel as a higher viscosity fluid. A change in temperature of 5C can cause a 0,1% change in the reported flow rate measurement.

A compensation equation is used to predict turbine meter behaviour based upon temperature and viscosity. The compensation equation uses dimensionless Reynold's numbers (Roshko and Strouhal) that take into account the expansion effects of materials based upon temperature. It is important to have at least three viscosities characterized for the use range to verify the turbine meter behaves as predicted. Unless requested, turbine meter behaviour is generally measured for only one viscosity.

With a 7:1 range ratio of low to high flow rate, the best measurement uncertainty of a turbine meter on a single viscosity is 0,27% of reading with a fluid and ambient air temperature stable within +/-2C.

Equations

A turbine measures All symbols used in equations will be defined in the Standard. Expanded measurement uncertainty is expressed using equation (1):

$$U_e = 2\sqrt{U_a^2 + U_{b1}^2 + U_{b2}^2 + \dots + U_{bn}^2} \quad (1)$$

The pulsation index is defined by equation (2):

$$I_p = \frac{\left(\overline{V_f}\right)_{\max} - \left(\overline{V_f}\right)_{\min}}{2\left(\overline{V_f}\right)_{\text{avg}}} = \frac{\left(q_v\right)_{\max} - \left(q_v\right)_{\min}}{2\left(q_v\right)_{\text{avg}}} \quad (2)$$

The flow-rate correction factor, based upon the pulsation index, is defined by equation (3):

$$F_p = \frac{\text{indicated_flow}}{\text{average_flow}} = \left(1 + 4ab_p I_p^2\right)^n \quad (3)$$

An axial flow meter compensation equation is based upon a Strouhal Number, equation (4), and a Roshko Number, equation (5).

$$S_r = \frac{f}{u} c^3 \quad (4)$$

$$R_o = S_r R_e = \frac{f L}{\nu} \quad (5)$$

A coriolis mass flow meter is described in equation (6):

$$\dot{M} = S_k \frac{A_c}{A_e} \left(\frac{1}{f_v}\right) \quad (6)$$

Electronics outputs and Uncertainty

Error Source	Type	Contribution	Distribution	Divisor	Contribution divided by Divisor	Squared contribution
Turbine Meter Repeatability (in a single viscosity)	A	0,05	Statistical	1	0,05	0,0025 U _a
Cal Lab Flow Stand – Flow Rate	B	0,05	Rectangular	1,732	0,029	0,0008 U _b
Cal Lab Flow Stand – Frequency	B	0,00	Rectangular	1,732	0,00	0,0000 U _b
Cal Lab Master Temperature Probe (+/- 28C)	B	0,00	Rectangular	1,732	0,00	0,0000 U _b
Viscosity Measurement (or table calculation)	B	0,25	Rectangular	1,732	0,144	0,0207 U _b
Density, no need to calculate from base density; under 7 bar, pressure effects are < 0.01% of reading	B	0,00	Rectangular	1,732	0,00	0,0000 U _b
Standard Error Estimate (SEE)	B					
Strouhal No Calculation	B					
Roshko No Calculation	B					
Interpolation Error	B					
Piping/Inlet Mismatch (difference between pipe inner diameter and unit inlet diameter)	B					

This section describes electronics used in addition to the primary mechanical/electrical transducer element and the extra contribution of error to the measurement uncertainty calculation.

The additional output electronics adds value to the primary element signal, such as: a) producing a scaled voltage from a frequency, b) compensating for temperature or viscosity or density, c) dampening flow pulsations, d) producing a communications port signal, to mention a few. The additional electronics can usually be remote-mounted because the value added electronics are typically more sensitive to extreme environmental conditions than the primary sensing technology.

Calibration Method and its Quality

This section describes the calibration method versus real use and includes discussions on a variety of topics that influence the quality of the unit calibration. The calibration shall be in an environment as close as possible to the intended use of the flow measurement instrument, as per the Laboratory Management Standard ISO 17025. The flow points and number of flow points shall be selected based upon the use of the flow measurement instrument. The inlet plumbing shall be the plumbing attached to flow measurement instrument when used. The length and configuration of inlet plumbing to the unit under calibration shall conform to the turbulence influences stated in the technology section of this standard, such as a length of 1, 10, 100 or 1000 times the pipe inner diameter. The fluid viscosity, density and temperature shall be within an acceptable tolerance range for the use of the flow measurement instrument. The fluid used in calibration may be substituted with fluids of like fluid properties in order to ensure the safety of the personnel performing the calibration and be compatible with the calibration system as well as reduce the cost of the calibration. A 4:1 ratio of calibration system versus unit under calibration shall be maintained, whenever possible, to ensure the integrity of the calibration measurement uncertainty maintains less than a 2% producer's risk of false acceptance. The fluid temperature shall be measured at the unit-under-calibration as well as the calibration references because mass flow and volumetric flow are not conserved. In some circumstances, the calibration must be performed in situation, such as flow measurement instruments used in pipelines for water or petroleum as well as in rivers and oceans.

Conclusion

Numerous industries and manufacturers have expressed interest in reviewing or contributing to this proposed standard. There is momentum to complete this technical document. For user convenience in the Internet world, the international standard shall be up-dated annually and accessible on-line by subscription as well as downloaded PDF file for sale.

Acknowledgements

Thanks for the encouragement to continue writing this standard to Alex Gahrahmat, Andy Vratny, Mike Roeder,

York Xu, Jon Heiner and, especially, Gerri Field, my wife.

References

- [1] W.G. Andrew, Applied Instrumentation in the Process Industries Volume 1: A Survey, Gulf Publishing Company, Texas, 1974.
- [2] R.W. Miller, Flow Measurement Engineering Handbook, McGraw-Hill, New York, 1996.
- [3] W. Guthrie and Hung-kung Liu, "Hands-on Workshop on Estimating and Reporting Measurement Uncertainty", NCSLI Conference, pp. 7-107, St. Paul, Minnesota, USA, March 2008.
- [4] ISO/TAG 4/WG 1, Guide to the Expression of Uncertainty in Measurement, International Standards Organization, Geneva, 1993.
- [5] L.O. Olsen, "Introduction to Liquid flow Metering and Calibration of Liquid Flowmeters", National Bureau of Standards Technical Note 831, NIST, USA, 1974.
- [6] ISO/IEC 17025:2005, General requirements for the competence of testing and calibration laboratories, International Standards Organization, Geneva, 2005.
- [7] Haefelfinger, Honegger, Gfeller, "A Novel Concept of a High Accuracy Calibration Rig", 14th Congress International de Metrologie, Paris, France, 22-25 Juin 2009.
- [8] Christopher David and Pierre Claudel, "Novel Water Flow Facility in France Range Extension to Low Flow Rates (10 000 ml/h down to 1 ml/h)", MAPAN - Journal of Metrology Society of India, Vol. 26, No. 3, 2011; pp. 203-209.
- [9] ANSI/NCSL Z540.3-2006, Requirements for the Calibration of Measuring and Test Equipment, ANSI, USA, 2006.