

# A calibrated physical flow standard for medical perfusion imaging

G. Kok<sup>1</sup>, N. Pelevic<sup>1</sup>, A. Chiribiri<sup>2</sup>, X. Milidonis<sup>2</sup>, M. Nazir<sup>2</sup>, M. Capstick<sup>3</sup>, S. Drost<sup>4</sup>,  
C. Poelma<sup>4</sup>, T. Schaeffter<sup>5</sup>

<sup>1</sup>VSL B.V., Thijssseweg 11, 2629 JA, Delft, Netherlands

<sup>2</sup>School of Biomedical Engineering and Imaging Sciences, King's College London, London, United Kingdom,

<sup>3</sup>ZMT Zurich MedTech AG, Zurich, Switzerland,

<sup>4</sup>Laboratory for Aero- and Hydrodynamics, Technische Universiteit Delft, Delft, Netherlands,

<sup>5</sup>Physikalisch-Technische Bundesanstalt, Berlin, Germany

E-mail (corresponding author): gkok@vsl.nl

---

## Abstract

In the medical sector, various imaging methodologies or *modalities* (e.g. MRI, PET, CT) are used to assess the health of various parts of the bodies of patients. One such investigation is the blood flow or *perfusion* of the heart muscle, expressed as the (blood) flow rate normalized by the mass of the volume of interest, with unit mL/min/g. Currently there is no physical flow standard for the assessment and validation of myocardial perfusion imaging methodologies, resulting in a large proportion of medical diagnoses being inaccurate and highly dependent on the scanner type, software used and the clinical operator. In the EMPIR 15HLT05 PerfusImaging project a phantom simulating myocardial perfusion has recently been developed with which imaging modalities can be tested. In this paper the construction and validation of the phantom is described which involved several iterations with design updates, computational fluid dynamics simulations, 3D printing of the phantom, ultrasound imaging velocimetry and magnetic resonance imaging (MRI). Dynamic contrast-enhanced MRI was performed to image the passage of a tracer through the phantom and estimate perfusion. Two flow models and associated data analysis methods to relate the measurement data with the reference flow rates are presented and discussed.

---

## 1. Introduction

In the medical sector, various imaging methodologies or *modalities* (e.g. MRI, PET, CT) are used to assess the health of various parts of the bodies of patients. One such investigation is the blood flow or *perfusion* of the heart muscle, expresses as the (blood) flow rate normalized (divided) by the mass of the volume of interest, with unit mL/min/g. A decreased perfusion of the heart muscle or myocardium is an imaging biomarker for increased risk for a heart attack and can cause chest pain.

Currently there is no physical flow standard for the assessment and validation of myocardial perfusion imaging methodologies, resulting in a large proportion of medical diagnoses being inaccurate and highly dependent on the scanner type, software used and the clinical operator. In the EMPIR 15HLT05 PerfusImaging project [1] a phantom simulating myocardial perfusion has recently been developed with which imaging modalities can be


tested. The design was based on an earlier, already existing phantom made by KCL [2].

In this paper the construction and validation of the phantom is described which involved several iterations with design updates, computational fluid dynamics (CFD) simulations, 3D printing of the phantom, ultrasound imaging velocimetry (UIV) and MRI measurements involving all partners. Dynamic contrast enhanced MRI (DCE-MRI) measurements were also performed. Various flow models to interpret the data will be discussed.


## 2. Phantom design

The phantom [3] mimics the complete flow through a human heart including the four heart chambers, pulmonary volume and the blood flow through the heart muscle. The part simulating the tissue of the heart muscle or the *myocardium* is of special interest in this study. The arteries in this tissue are simulated by more than 200 parallel channels or capillaries with cross-sectional area ranging from 1

to 7 mm<sup>2</sup>. Upstream of the channels is a pre-chamber with a volume of 36 mL. This part was produced by 3D printing and is shown in Figure 1. The engineering and production of the phantom was done by the company Zurich MedTech (ZMT). In Figure 2 a schematic drawing of the phantom is shown including the location of the MRI imaging plane. The flow rates through the aorta (cardiac output) can be varied between 2 and 5 L/min and the perfusion rate through the myocardium between 1 and 5 mL/min/g, corresponding to flow rates between 85.3 and 426.5 mL/min. In this study, the flow rate through the aorta was set around 4 L/min and the full range of flow rates through the two myocardia were tested.



**Figure 1:** Photograph of the 3D printed myocardium indicating the main compartments.




**Figure 2:** Schematic overview of the phantom. The MRI imaging locations MRI-A and MRI-B are in reality in the same plane (dashed lines). The four heart chambers are denoted by RA, RV, LA and LV (right atrium, right ventricle, left atrium, left ventricle).

### 2.1 CFD simulations

The design goal was a channel configuration with a rotationally symmetric flow profile with maximum flow velocity at the centre and linearly decreasing flow velocity to -30 % at the outer wall (for channels with identical cross sections). While Poiseuille's law for a laminar flow regime suggests a linear increase of +30 % in channel length towards the outer wall to obtain the desired linear variation in flow velocity, the OpenFOAM CFD solver was used by VSL to confirm this expectation and to ensure that the flow FLOMEKO 2019, Lisbon, Portugal

regime would be rotationally symmetric, at least in the model. The final design had four inlets from the sides to the pre-chamber in which a laminar flow element upstream of the channels was installed to remove any form of potential pressure gradient and turbulence in the cross section (the Reynolds number is below 250 in the pre-chamber). In Figure 3 two plots of the results of the CFD simulation are shown. Approximately 4 million cells were used. For the large octagonal channels the CFD results are in line with the simple analytical Poiseuille model, see Table 1. For the smaller square channels with side lengths 1.0, 1.2, 1.4 and 1.8 mm, the predicted flow rates by CFD are smaller than the analytical ones, which is probably due to the mesh size of the CFD simulation.




**Figure 3:** Some results of the velocity field calculated by CFD. Top: cross-cut along the main axis of the phantom. Bottom: cross cut perpendicular to the axis at the mid-myocardial level. The black box indicates the octagonal channels used in Table 1.

**Table 1:** Flow rate per channel as a fraction of the flow in the octagonal channels in the box of Figure 3 – bottom. Numbering is from top left to bottom right.

Channel nr.	1	2	3	4	5	6	7	8
Percentage	70%	78%	89%	100%	100%	89%	78%	70%

## 2.2 PIV measurements

In order to measure the actual flow rates through the channels of the phantom, measurements with ultrasound imaging velocimetry (UIV) were performed by TU Delft. The earlier version of the phantom was based on thin walled straws [2]. The current version was 3D printed and had thicker walls [3], which turned out to be problematic for the UIV measurements. Some indicative results were finally obtained suggesting linearly decreasing flow velocities with radius, but with an uncertainty higher than the target uncertainty of 10 %. In Figure 4 a schematic view of the measurement set-up and a measurement image are shown.




**Figure 4:** Top: Schematic drawing of the ultrasound imaging velocimetry (modified) set-up, showing the compartment in light blue and the ultrasound transducer in grey. FOV indicates field of view. Bottom: tracer visualization at the outflow. By correlating the intensity within the two red squares in the capillary in the centre, the mean flow can be determined.

## 2.3 Phase contrast MRI measurements

As alternative to the UIV measurements phase contrast (PC) MRI measurements were performed with an estimated uncertainty of 10 % for the flow velocities. Five flow rates between 85.3 and 426.5 mL/min were each measured twice, and the results were analysed in various ways by KCL and VSL. In Figure 5 one such analysis by VSL is shown. The channel geometry is slightly different than that of Figure 3, but the expected decrease in flow rate is the same. The octagonal channels on a horizontal line were identified by an algorithm, and the flow rates through channels were calculated by integrating the velocity field and expressed as a percentage relative to the flow rate in channel 5. A decrease of  $30 \pm 10$  % with respect to the maximum

flow rate is visible for all flow rates, be it not completely symmetrical between left and right side.



**Figure 5:** Top: PC-MRI measurement image with identified octagonal channels on a horizontal line used for integration of the measured velocity field (except for the most left one). Bottom: Relative flow rate distribution over the channels at five flow rates as measured by PC-MR, normalized to the flow rate in channel nr. 5.

## 3. DCE-MRI and flow modelling

### 3.1 Dynamic Contrast Enhanced MRI protocol


For clinical measurements of myocardial perfusion in patients a so-called Dynamic Contrast Enhanced MRI protocol is being used. A bolus of contrast agent (CA, e.g. gadolinium) is injected in the patient upstream of the heart and this is simulated by injecting such a bolus at a specific point in the phantom representing the vena cava. The CA concentration in the blood is roughly proportional to the measured MRI signal. The bolus of CA mixes with the blood (or water) and passes through the four heart chambers. Then, the MRI signal is measured as function of time in (a pipe simulating) the aorta, upstream of the myocardium. The resulting time series is called the arterial input function (AIF). In the same image, the concentration inside the myocardium is also visible. After conversion from signal intensity to CA concentration, the perfusion of the myocardium can be estimated. In patients typically three slices of the heart are being measured, though one slice at a time is used during the analysis phase. In the experimental work with the phantom one slice of the

phantom has been measured and analysed. To reduce the effect of the non-linearity of the relationship between CA concentration and MRI signal intensity a dual bolus scheme is typically used. In this scheme a first ‘pre-bolus’ diluted in saline at a ratio 1:9 is injected from which the AIF is derived and scaled by a factor 10. Subsequently, a second undiluted bolus is injected from which the myocardium signal is taken. As the CA concentration in the myocardium is much lower (and more spread out over time) than in the aorta, the MRI signal magnitudes of AIF and myocardium measurements are similar, and the effects of non-linearity and signal saturation are diminished.

In the next sections various ways of analysing the data are presented.

### 3.2 Standard theory

The standard theory and method for quantification of the perfusion assumes a system of volume  $V_{sys}$  with one inlet and one outlet, see Figure 6. The AIF is measured at the inlet by MRI-A (aorta) and is proportional to the CA concentration  $c_{in}(t)$ . The average CA concentration in  $V_{sys}$  at time  $t$  equals  $c_{sys}(t)$  and is measured by MRI-B, the measurement of the myocardium. The stationary flow rate through the system equals  $q_{in} = q_{out}$  and the quantity of interest is the perfusion  $f = q_{in}/V_{sys}$ .



**Figure 6:** Standard system view used in perfusion quantification with DCE-MRI.

Based on a mass balance of CA in the system, the following equation can be derived

$$V_{sys}c_{sys}(t) = \int_0^t q_{in} c_{in}(s) ds - \int_0^t q_{out} c_{out}(s) ds \quad (1)$$

Assuming a linear and stationary system with impulse response function  $h(t)$ ,  $c_{in}(t) = c_{sys}(t) = 0$  for  $t < 0$  s and using  $q_{in} = q_{out}$  the outlet concentration can be written as

$$c_{out}(t) = \int_0^t c_{in}(t-s)h(s)ds = (c_{in} * h)(t) \quad (2)$$

where \* denotes the convolution operation. Defining  $R_f(t) = q_{in}/V_{sys} \left(1 - \int_0^t h(s) ds\right)$  and performing some mathematical transformations one can derive that

$$c_{sys}(t) = (c_{in} * R_f)(t). \quad (3)$$


Solving convolution Equation (3) for  $R_f(t)$  the perfusion  $f$  follows from  $f = R_f(0) = \max(R_f(t))$ .

In practice a correction factor has to be applied for the accessible part of the myocardial tissue for the CA (‘volume fraction’), or the ‘non-plastic fraction’ in the case of a phantom.

Advantages of the standard approach are its relative simplicity and the fact that the volume  $V_{sys}$  doesn’t need to be known. Possible limitations are that the measured concentration by MRI-B is in practice (at least for the phantom) rather an outflow CA concentration  $c_{out}(t)$  than the system average concentration  $c_{sys}(t)$ , and interaction between different parts of the myocardium is not modelled.

### 3.3 Alternative model

An alternative model and method for calculating the perfusion is based on Figure 7, where a system with a common pre-chamber is shown and two subsequent compartments in parallel. Note that this model is not limited to three compartments, but can be generalized to an arbitrary number of compartments. For ease of presentation only three compartments are used in this paper.



**Figure 7:** Alternative system view for perfusion quantification with DCE-MRI. The compartment volumes are denoted by  $V_i$ , compartment average concentrations as function of time are denoted by  $c_{sys,i}(t)$ , compartment outflow concentrations are denoted by  $c_i(t)$  and outflow flow rates by  $q_i$  for  $i = 0, 1, 2$ .

It is again assumed that at location MRI-A the MRI signal is proportional to the CA inlet concentration  $c_{in}(t)$ . However, at MRI-B<sub>1</sub> and MRI-B<sub>2</sub> the signal is assumed to be proportional to the respective outlet concentrations. The perfusion  $f_i$  in compartment  $i$  is defined by  $f_i = q_i/V_i$ , the impulse response function for compartment  $i$  is denoted by  $h_i(t)$  and the mean transit time  $T_i$  of a unit of CA through a compartment  $i$  is given by  $T_i = \int_0^\infty t h_i(t) dt$ . From system inlet to outlet  $i = 1$  or  $2$  the mean transit time  $T_{0i}$  is given by  $T_{0i} = T_0 + T_i$  and the impulse response function  $h_{0i}$

is given by  $h_{0i} = h_0 * h_i$ . The relationships of interest for the measured MRI signals are

$$c_i(t) = (c_{in} * h_{0i})(t), \quad i = 1, 2 \quad (4)$$

which is similar to Equation (3).

In order to be able to calculate the perfusions an additional hypothesis is needed. The assumption made is that there exists a constant tissue delay factor  $d$  for which following relation holds true for any compartment  $i$ :

$$T_i = d V_i / q_i \quad (5)$$

From Equation (5) then follows that  $f_i = d / T_i$ . Using  $q_{in} = q_1 + q_2$  all perfusions as well as flow rates can be calculated when the measurement results  $c_{in}(t)$  and  $c_i(t)$  are available.


When a dual bolus scheme is being used or when there is a significant delay before the CA arrives at the myocardium additionally an additional time offset  $\tau$  has to be used in Equation (5), leading to  $T_i = \tau + d V_i / q_i$ . Note that this is also the case for the standard method. In the assessment of the methods discarding  $\tau$  and only using the main bolus gave better results and these are shown in Table 2.

Advantages of this alternative method are that it might be more realistic and thus might yield more accurate results, as it models the measurements at MRI-B<sub>i</sub> as proportional to the compartment outflow concentration rather than the compartment's average concentration, and it models a common large blood vessel (or phantom pre-chamber) upstream of smaller myocardium blood vessels (or phantom channels) in parallel. Disadvantages are the required knowledge of the (relative) volumes of the compartments (which is possible for a phantom but very difficult for tissue of patients) and the hypothesis of constant tissue delay factor (of which the exact value is required if absolute values of perfusion are of interest). Another disadvantage is that determination of the mean transit time requires integration over the complete relevant time domain of the signals  $c_{in}(t)$  and  $c_i(t)$ , which is possible for a phantom, but problematic for patients where CA may re-enter the aorta and myocardium resulting in an undesired 'second-pass' perfusion MRI signal. Another assumption is that  $c_0(t)$  is homogeneous at the outlet plane of the initial compartment. (A similar assumption is used in the standard method.)

### 3.4 Comparison of data analysis methods

The two analysis methods described in section 3.2 and 3.3 were applied to DCE-MRI data acquired for FLOMEKO 2019, Lisbon, Portugal

the phantom described in section 2. The target quantity was the ratio  $r$  of perfusion through the outer ring to the perfusion through the inner circle, i.e.  $r = f_{out} / f_{in}$ , see also Figure 8.



**Figure 8:** Left: Schematic partition of image in two parts of equal area. Right: Phantom with local pixel-wise perfusion values based on inverse transit times and assuming  $d = 1$ .

The two segments have equal area. The channels close to the outer boundary were taken out in this analysis, as measurement results were unreliable in this area due to partial volume effects.


The results of the calculations for both methods are shown in Table 2. In this case both methods turn out to perform equally well and are both in line with the reference value based on PC-MRI measurements (which are in-line with CFD and analytical calculations).

**Table 2:** Results of the calculations using the standard method ( $r^A$ ) and the alternative method ( $r^B$ ). The reference value based on PC-MRI measurements is  $r^A = 0.87 \pm 0.05$ .

$q_0^{ref} / (\text{mL/min})$	$r^A$	$r^B$
55	0.89	0.79
110	0.79	0.82
165	0.85	0.87
220	0.91	0.92
275	0.82	0.85

## 4. Discussion

The two analysis methods presented in this paper both rely on the validity of some additional assumptions. One of these assumptions is that the CA concentration at the inlets of all channels are equal at the same point in time and that the images are rotationally symmetric. Inspection of the raw MRI images directly shows that this is not the case, see Figure 9. Especially at the bottom the MRI signal appears to be higher.



**Figure 9:** Raw MRI image of the phantom myocardium at flow rate 170 mL/min, showing imperfect rotational symmetry.

As the contrast agent has a higher density than the water some stratification seems to happen especially in the case of the lowest flow rates. Updated designs with passive mixers may improve the situation. Another practical issue that was encountered was that removing all the air bubbles from the phantom was not so straight forward.

The comparison of the performance of the methods may be obfuscated by effects such as stratification effects as discussed in the last paragraph. In the analysis algorithm various parameters need to be selected, such as an image base line value, the pixels to be included in the analysis and similarly for the selection of time points. Each of these choices has an effect on the results which in some cases is non-negligible. These issues are also encountered in medical practice and the phantom can help to assess the importance of each parameter.

The goal of the phantom is to make MRI perfusion measurements more quantitative and more comparable across different MRI manufacturers, software settings and operators when measuring real patients. Even if the alternative method would turn out to perform better in a next version of the phantom, its requirements in terms of necessary (relative) volume information and measurement time range and other assumptions may limit its applicability in a medical setting. Application to sample patient data is a necessary next step to assess the usefulness of the alternative method and this will be a next step in our research.

## 5. Conclusion

In this paper a calibrated physical flow standard for medical perfusion imaging was presented, i.e. a phantom mimicking a human heart with known flow rates through its channels which can be used to assess the performance of MRI scanners as well as of PET and CT scanners.

Reference values for the flow rates through the channels were measured with phase contrast MRI with an uncertainty of approximately 10 % and were in line with CFD and analytical calculations. PIV measurement results were only indicative due to the thickness of the 3D-printed material. Overall flow rates are being measured with flow meters.

Two models and associated analysis methods were presented that can be used to relate dynamic contrast enhanced MRI measurements with the reference flow rates. The methods performed approximately equally well for the phantom measurement data.

Current research in the project consists of measuring the phantom with different scanners, extending the phantom to a multi-compartment phantom simulating more faithfully human tissue, and applying the alternative data analysis method to real patient data in order to assess its usefulness in a medical setting.

## Acknowledgment

The authors gratefully acknowledge funding of this research by the EMPIR 15HLT05 PerfusImaging project. The EMPIR initiative is co-funded by the European Union's Horizon 2020 research and innovation programme and the EMPIR Participating States.

## References

- [1] EMPIR project 15HLT05 PerfusImaging, Metrology for multi-modality imaging of impaired tissue perfusion, <https://www.ptb.de/empir2018/perfusimaging/home/>
- [2] Chiribiri et al., "Perfusion Phantom: An Efficient and Reproducible Method to Simulate Myocardial First-Pass Perfusion Measurements With Cardiovascular Magnetic Resonance", *Magnetic Resonance in Medicine*, 2012, DOI 10.1002/mrm.24299
- [3] Milidonis et al., *Dynamic contrast-enhanced and phase contrast MRI of a novel 3D printed cardiac phantom mimicking transmural myocardial perfusion gradients*, ISMRM conference, Paris 2018.