

## DOUBLE CHANNEL DYNAMIC ERROR CORRECTION OF METHANE SENSOR

*Ryszard Bogacz*<sup>1</sup>, *Beata Krupanek*<sup>2</sup>

<sup>1</sup> Silesian University of Technology, Faculty of Electrical Engineering Institute of Measurement Science, Electronics and Control, Gliwice, Poland, pl email: ryszard.bogacz@polsl.pl

<sup>2</sup> Silesian University of Technology, Faculty of Electrical Engineering Institute of Measurement Science, Electronics and Control, Gliwice, Poland, email: beata.krupanek@polsl.pl

**Abstract** – The primary reason for measuring methane is to improve the safety of the mines. In recent years, there have been many fatalities in underground coal mine explosions in which methane was a contributing factor. The rapid detection of methane is very important from the point of view of safety of mine workers. This paper presents a concept of fast methane detection by prediction its concentration in dynamic states. Two methods of dynamic corrections of methane concentration - single-channel and double-channels were described.

**Keywords:** dynamic error correction, methane sensor, pellistor sensor, safety in mines

### 1. INTRODUCTION

Coal mine methane is a term given to the methane gas produced or emitted in association with coal mining activities either from the coal seam itself or from other gassy formations underground. In range between 5% and 15%, known as the explosive range, methane can be ignited easily with the presence of an ignition source to create a violent methane explosion that may propagate in the presence of combustible coal dust [22]. Coal mine methane has always been considered as a danger for underground coal mining as it can create a serious threat to mining safety and productivity due to its explosion risk [19].

One of the most important duties of ventilation in underground coal mines is to keep methane levels well below the explosive limit by diluting methane emissions that occur during mining. From these reasons the methane concentration should be controlled by continuous measurements [10, 14].

Methane sensors use different detection technologies according to the three gas categories and these are pellistors, electrochemical, and infrared devices [3]. Typically pellistors are used for detecting methane and any flammable gas [6, 21]. One of the disadvantages of pellistor sensor is a long response time [11, 15]. This defect can be significantly reduced by using a dynamic correction method [16,17,18]. Such a method was shown in [1, 2]. Article shows a way for use the correction method for pellistor sensors in coal mines for reducing the time to detect methane.

### 2. DYNAMIC ERROR OF METHANE SENSOR

Pellistors are gas sensors that detect combustible gases and vapours in air (or atmospheres containing oxygen), in concentrations approaching the explosive range. In this case, they are operating in a catalytic mode where the target gas is burnt and the heat liberated is measured by the sensor. The working principle of catalytic sensors is based on flammable gas oxidation: when a combustible gas comes in contact with the catalyst surface it is oxidised. The reaction releases heat, which causes the resistance of the wire to change.

A catalytic palletised resistor (or "Pellistor") consists of a very fine coil of platinum wire, embedded within a ceramic pellet. The signal produced by measurement setup is proportional to the gas concentration up to the lower explosive limit, or LEL [8]. The sensor will detect all combustible gases and vapours, although the response to higher hydrocarbons may not be high [12].

The thermal conductivity sensor comprises two beads, one of which is exposed to the target gas (the Detector) and the other sealed inside a chamber containing air (the Compensator). When the detector bead is exposed to a gas whose thermal conductivity is significantly different to that of air, the rate of heat loss from the bead will change, as will its resistance. This change measured is compared with the compensator bead [6].

The response time of sensor depends on the gas being detected: the greatest is the gas weight or the molecular dimension the longer will be the response time. A  $T_{90}$  time response to methane detection is usually provided in 5 - 10 seconds [2, 12, 22]; a slight increase occurs when the sensor is protected by a sintered filter [12, 14].

Because of catalytic combustion the pellistor sensor must be placed behind a flameproof enclosure [7]. It normally contains a sintered metal flame arrester that also acts as a primary diffusion barrier through which the gas must pass before reaching the sensing element [5, 4, 8].

Having passed through this initial barrier, there is a second barrier at the entrance to the sensor can before the gas reaches the sensor beads. All of this factors cause the much longer time to receive the stable value of a methane concentration in the case of its rapid growth [20].

### 3. SINGLE CHANNEL DYNAMIC CORRECTION

Dynamic errors are primarily caused by system delays. In this case dynamic error is caused by delay introducing by flameproof arrester. This delay can be variable because the shield is placed in dusty and polluted places in mine. Changes of time constant may be large what illustrates the figure 1.




Fig. 1. Changes of pellistor sensor time constant

Determination of a dynamic correction algorithm of a real measuring sensor requires knowledge of its dynamics especially measurement system time constant needed for formulation of differential equations. Therefore, correction algorithms must always be determined for specific measuring transducers and measurement systems, and the resultant equations must only be applied to the particular sensors [1, 2].

Dynamic properties of methane sensor can be modeled by a linear differential equation of n-th order, but usually transducer can be described by first order equation. This expression describes the dynamic processing error. When measured quantity changes occurs respectively slow, the parts of expression containing derivatives take values close to zero and this equation takes the form of the equation of ideal processing. Sensor output can be represented as the sum of the input quantity and dynamic error. This relationship allows us to determine the dynamic error at the output of the transducer to any changes of known input quantity [2, 13].

Dynamic correction is usually realized by microcontroller due to its numerical complexity. We assume that the measured value is converted to other quantity realised by some function. Dynamic correction is based on solving the inverse function to that which is taken from measurement of the quantity, which means that the output values are equal to the values of the measured quantity. It could be said that the measuring circuit performs tasks of ideal transducer. Assuming that, the measures are performed by AD converter the differential equations takes the discrete form. Now, because of correction occurs in discrete moments only the instantaneous values of the transducer output are predicted.

In general case the described algorithm has a recurrent solution which was described in [2]. First order model is a special case, for which there is no form of recursive algorithm correction and the algorithm has simply form.

The experimental results of methane dynamic correction for the sudden change of the concentration of methane was shown in paper [1, 13]. It can be seen that the algorithm of correction works properly with a delay equal to two sampling periods. This accelerate the measurement almost several times.

Method described in paper [1, 2] lead to the conclusion that it is possible the correct dynamic error of penetration of methane by flameproof arrester almost real time, basing only on the instantaneous values of methane concentration inside the pellistor sensor enclosure. Correction is performed on the basis of dynamic correction of the discrete changes of the concentration outside the enclosure, and the result is a significant decrease in time to obtain a result. This method can significantly helps to improvement the work safety of miners.

### 4. DOUBLE CHANNEL DYNAMIC CORRECTION

In a situation when the dynamic properties of sintered shield changes over time, what happens as a result of the contamination, the use of single channel correction model causes additional errors resulting from the difference between the adopted for the calculation and the actual value of time constant. In this situation, it is necessary to identification the time constant. The most useful is a method that does not require the use of calibration mixtures, called "blind correction" [16, 17, 18] or double channel correction. The essence of that method involves the parallel operation of two measuring heads with different time constants which values are not known. Software processing algorithm consists in this case of two steps: first self-identification of time constant (or both), and next a correction like that for single channel method. The structure of the measuring system implementing double channel correction is shown in Fig. 2.




Fig. 2. Processing system for double channel correction.

Measured signal is brought to the inputs of two analog measuring transducers. Both of them have the same type of dynamic models, but differ in the model coefficients. It was assumed that these are the first order models with different time constants. The actual values of the output signals are sampled with the same frequency for both transducers. The samples are converted by AD converters. Obtained samples are processed by identification algorithm to designate the model coefficients. Then it is realized a correction algorithm similar to single channel correction based on the obtained time constant.

Model used in single channel correction need only time constant identification of 1-st order differential equation. In this case:

$$x(t_i) = y_1(t_i) + T_1 \cdot \dot{y}_1(t_i) = y_2(t_i) + T_2 \cdot \dot{y}_2(t_i) \quad (1)$$

where  $t_i, i = 1, 2, \dots$  is a moment of signal measure by both ADCs with frequency  $f_s = T_s^{-1}$ ,  $T_s$  is sampling period. The initial conditions  $y_1(0)$  and  $y_2(0)$  should be taken into account. Using the results from  $N$  successive sampling moments obtained a set of equations:

$$\begin{cases} y_1(t_1) + T_1 \cdot \dot{y}_1(t_1) = y_2(t_1) + T_2 \cdot \dot{y}_2(t_1) \\ y_1(t_2) + T_1 \cdot \dot{y}_1(t_2) = y_2(t_2) + T_2 \cdot \dot{y}_2(t_2) \\ \vdots \\ y_1(t_i) + T_1 \cdot \dot{y}_1(t_i) = y_2(t_i) + T_2 \cdot \dot{y}_2(t_i) \\ \vdots \\ y_1(t_N) + T_1 \cdot \dot{y}_1(t_N) = y_2(t_N) + T_2 \cdot \dot{y}_2(t_N) \end{cases} \quad (2)$$

It allows to determine the time constants  $T_1$  and  $T_2$  of both measuring transducers. This set of equations must be solved in order to obtain information about the constant  $T_1$  and  $T_2$  in a way that minimizes the errors of identification [18].

In case of methane concentration measurements the both measuring channel has the different time constants but they are described by the same equations. For each of the two measuring circuits shown in Fig. 2 concentration outside the flameproof enclosure are equal:

$$C_{o1}(t) = C_{o2}(t) = C_o(t), \quad (3)$$

where  $C_{o1}(t)$  is the concentration of methane out of the first enclosure, and  $C_{o2}(t)$  is the concentration of methane out of the second enclosure. Taking into account (1) and (3) it can be written:

$$T_{o1} \frac{dC_{i1}(t)}{dt} + C_{i1}(t) = T_{o2} \frac{dC_{i2}(t)}{dt} + C_{i2}(t), \quad (4)$$

where  $T_{o1}$  i  $T_{o2}$  are the time constants of flameproof shields respectively the first and second channel, and where  $C_{i1}(t)$  is the concentration of methane in the first enclosure, and  $C_{i2}(t)$  is the concentration of methane in the second enclosure. For any two time moments  $t_1$  and  $t_2$  it can be written:

$$\begin{cases} T_{o1} \frac{dC_{i1}(t_1)}{dt} + C_{i1}(t_1) = T_{o2} \frac{dC_{i2}(t_1)}{dt} + C_{i2}(t_1) \\ T_{o1} \frac{dC_{i1}(t_2)}{dt} + C_{i1}(t_2) = T_{o2} \frac{dC_{i2}(t_2)}{dt} + C_{i2}(t_2) \end{cases} \quad (5)$$

Assuming a constant sampling period of methane concentration inside the shield equal to  $T_d$ , numbering of sampling moments can be made by the variable  $k$ . The methane concentration predicted for the same sampling moments should be identical in both measuring channels for every time instant  $k$ :

$$\hat{C}_{o1}(k) = \hat{C}_{o2}(k) = \hat{C}_o(k), \quad (6)$$

Equation (6) determine the predicted values of the measured values (concentration of methane out of the enclosure) in both measuring channels.

Similar like for the single channel correction, the methane concentration can be predicted for each independently measurement channel for any  $k$  time moment. For the first measurement channel, methane concentration is determined by the formula:

$$\hat{C}_{o1}(k) = A_{k+1,1} \tilde{C}_{i1}(k+1) + A_{k,1} \tilde{C}_{i1}(k), \quad (7)$$

and for the second channel:

$$\hat{C}_{o2}(k) = A_{k+1,2} \tilde{C}_{i2}(k+1) + A_{k,2} \tilde{C}_{i2}(k). \quad (8)$$

The values of the coefficients  $A_{k+1,1}$  and  $A_{k,1}$  of equation (7) are calculated from formulas:

$$A_{k+1,1} = \frac{1}{1 - \varphi_1}, \quad A_{k,1} = \frac{-\varphi_1}{1 - \varphi_1}, \quad (9)$$

and values of the coefficients  $A_{k+1,2}$  and  $A_{k,2}$  of equation (8) are calculated from formulas:

$$A_{k+1,2} = \frac{1}{1 - \varphi_2}, \quad A_{k,2} = \frac{-\varphi_2}{1 - \varphi_2}, \quad (10)$$

where

$$\varphi_1 = e^{-\frac{T_s}{T_{o1}}}, \quad \varphi_2 = e^{-\frac{T_s}{T_{o2}}}. \quad (11)$$

In order to minimize prediction errors as the final result assumed the average of results from two channels in  $k$  moment. Next the predicted value was calculated from several subsequent results.

## 5. RESULTS OF CORRECTION

Experimental results of methane dynamic correction for the sudden change in the concentration of methane was shown in Fig. 3. Although a small uncertainty of the measurement data there are large errors in the predicted values which is due to the strengthening effect of random errors by the dynamic correction algorithms [9]. Simple averaging algorithm, reduces the final error dynamic correction by nearly half. The time required to obtain a measurement result shortened to less than one second. It is the time required to obtain a three consecutive reconstructed results for the filtration algorithm.




Fig. 3. Prediction methane concentration by double channel correction method.

Table 1. The calculated times  $T_{90}$  and maximum error values  $\Delta_m$ .

	$T_{90}$ , s	$\Delta_m$ , %
The answer to the step change in first channel	4,8	86
The answer to the step change in second channel	9,3	91
Results of prediction in first channel	0,9	8
Results of prediction in second channel	0,9	38
Result after averaging from both channels	0,9	20
Results after filtration	1,5	10

Most important parameter for prediction is the time of receipt of the final result with the required accuracy. The study shows that the developed algorithms provide significant reduction of time to obtain the result. For sample period equal 0.3 s, time of obtain the results is about 1 s for the single channel correction and simple filtration algorithms. In the case of double channel correction this time increases due to the fact that the prediction must be preceded by time constant identification.

## 6. CONCLUSIONS

Described correction method may be used to reduce the time for obtaining the result of measurement of the methane concentration after the step change in the mining conditions. Correction is performed on the basis of dynamic correction of the discrete changes of the concentration outside the enclosure, and the result is a significant decrease in time to obtain a result. In the coal mines pellistor shields might soil, which leads to changes in the dynamics of methane diffusion. Double channel correction method can be used in a situation where the time constant of the model is changing or is not known. It is ideal in mine operating conditions. In such a case requires the use of two measuring heads with shields with different time constants. The methods can be used on a almost in real time, which means that the execution time of correction algorithms is so short that it does not introduce a significant delay in obtaining the final result.

Method described in this paper lead to the conclusion that it is possible the current dynamic error correction of penetration of methane by flameproof arrester, based on the instantaneous values of methane concentration inside the enclosure measured using pellistor sensor. This method can significantly helps to improvement the work safety of miners.

## REFERENCES

- [1] R. Bogacz, B. Krupanek, "Dynamic Error Correction of Methane Sensor", *Int Journal of Electronics and Telecommunications*, 2014, v.60, no.4, pp. 287–289.
- [2] R. Bogacz, B. Krupanek, "Selected methods for acceleration the measurements of gas concentration in transient states", *Pomiary Automatyka Kontrola*, 1/2014, pp.26-30 (in polish).
- [3] J. Chou, "Hazardous Gas Monitors", A Practical Guide to Selection, Operation and Applications, *McGraw-Hill Book Company*, New York, 2000.
- [4] K. H. Clifford, S. W. Webb, "Gas transport in porous media", *Springer*, 2006.
- [5] J. Crank, "The mathematics of diffusion", *Oxford University Press*, New York, 1993.
- [6] J. Fraden, "Handbook of modern sensors: physics, designs, and applications", Jacob Fraden–3rd ed., *Springer*, 2004.
- [7] S. Grossel, "Deflagration and Detonation Arresters", *Center for Chemical Process Safety/AIChE*, 2002.
- [8] R. E. Henderson, "Understanding Combustible Sensor Performance", *International Fire Protection 2006* vol.27 pp.55-59.
- [9] J. Jakubiec, "A new conception of measurement uncertainty calculation", *Acta Physica Polonica A*, vol.124, 2013, pp.436-444.
- [10] F. N. Kissell, "Handbook for Methane Control in Mining", *Department of Health and Human Services, Information Circular*, 2006.
- [11] S. D. Kolev, M. Adam, I. Barsony, A. van den Berg, C. Cobianu, "Mathematical modelling of porous silicon based pellistor-type catalytic flammable gas sensors", *Microelectronics Journal* 29/1998, pp.235–239.
- [12] M. Krawczyk, J. Namiesnik, "Application of a catalytic combustion sensor (pellistor) for the monitoring of the explosiveness of a hydrogen-air mixture in the upper explosive limit range", *Journal of Automated Methods & Management in Chemistry*, vol.25, no.5, pp.115-122.
- [13] B. Krupanek, R. Bogacz, "Correction of dynamics errors of measurements the methane concentration", 13th IMEKO TC10 Workshop on Technical Diagnostics, *Advanced measurement tools in technical diagnostics for systems' reliability and safety*, June 26-27, 2014, Warsaw, Poland, pp. 157-161.
- [14] A. Kumar et all, "Application of gas monitoring sensors in underground coal mines and hazardous areas", *International Journal of Computer Technology and Electronics Engineering*, vol.3, Iss.3, 2013, pp.9-23.
- [15] G. McRobbie, F. Clark, C. Tandy, "The Performance of a Flammable Gas Sensing Pellistor Bead With Respect to the Material Properties of the Support Arms", *International ANSYS Conference*, 2002.
- [16] J. Nabelec, J. Nalepa, "The 'Blind' Method of Dynamic Error Correction for Second Order System", *XVII IMEKO World Congress*, 2003, pp.841-846.
- [17] J. Nabelec, J. Nalepa, "A Simulation Investigation of Differential Algorithm for the "Blind Correction" of Dynamic Error in Measuring Channels with Periodic Nonstationarity", *XVIII IMEKO World Congress*, 2006.
- [18] J. Nalepa, "Correction of Dynamic Error by the „Blind” Method. A Differential Algorithm Simulation Study”, *10th IMEKO TC7 Int. Symp. on Advances of Measurement Science*, 2004, pp.109-114.
- [19] K. Noack, "Control of gas emissions in underground coal mines", *International Journal of Coal Geology* 35, 1998, pp.57–82.
- [20] G. Rose, I. Zdanevitch, "A New Method Using a Catalytic Sensor for the Identification and Concentration Measurement of Combustible Gases", *Sensors and Actuators B*, vol. 24-25, 1995, pp.426-428.
- [21] G. Rose, "Microsensors for Methane and Other Combustible Gases", *Euroensors XI, The 11-th European Conference on Solid State Transducers*, Warsaw, Poland, SEPT 21-24, 1997, pp.123-126.
- [22] C. D. Taylor, J. E. Chilton, G. V. R. Goodman, "Guidelines for the Control and Monitoring of Methane Gas on Continuous Mining Operations", *Department of Health and Human Services, Information Circular*, 2010.